TFF-Micro: Tangential flow filter for separation of medium/large EVs (>150 nm)

amsbio

About TFF-Micro.

TFF-Micro is a filter cartridge containing polysulfone hollow fibers

(200nm pores), which allows the separation of medium/large EVs from small EVs and other solutes in different fluids (cell conditioned media, urine, plasma, serum etc..).

Small EVs, proteins and other soluble molecules pass through the hollow fiber pores (permeate), whereas medium/large EVs are concentrated in the retentate. Large EVs can be easily recovered with a syringe from the filter cartridge.



Sterility: The TFF-Micro is provided sterile. Once used the filter can besterilised by Beta irradiation. Do not autoclave the TFF-MV.

Plasma/serum

Storage.

Store the device at room temperature.

EV filtration procedure.

1- Sample injection by filtration (optional).



- Remove the screw from the sample injection nozzle (position 1).

5-20 ml

5-10 times

- Insert a syringe filter in the nozzle. The device is compatible with multiple filter types. The size of the filter to be used is user dependent.



2- Tangential Flow Filtration.



- Pick up the sample with syringe A and insert on the top of the filter.

- Insert a clean empty syringe in the Tangential flow filtration nozzle (position 2).

- Rotate the valve 3 to the position indicated by arrows in figure 3.

- Inject the sample into the filter by pushing the piston of syringe A. The sample passes through the filter, filling the syringe B.

- Rotate the valve 3 to the position indicated in figure 4.

- Open the permeate nozzle 4b by removing the cap.

- Insert a new clean syringe (C) in position 5.

- Set a permeate collection tube under the nozzle in position 4b, as indicated in figure 4.



- Start the filtration process by pushing the syringes B and C alternatively upwards and downwards. The permeate starts to flow to the collection tube, while the medium/large EVs are getting separated in retentate.

- Continue the filtration process until syringes are completely empty and there is no more permeate running.

amsbio





- While closing the flow through valve in position 3, disconnect the syringe B and fill it with PBS (\sim 10ml).

- Reattach the syringe B, enable the flow through valve in position 3 and wash the filter with PBS until the syringes are completely empty in order to remove residual smaller particles (figure 5). Repeat the step once.

4- Recovery of medium/large EVs.



- In order to collect retained medium/large EVs, close the permeate nozzle in position 4b.

- Rotate the valve in position 6 as indicated in figure 6 and disconnect the syringe C.

- Load syringe C with small amount of PBS (2ml).



Reattach syringe C to nozzle in position 5 and enable the flow through filter by rotating valve in position 6 as indicated in figure 7.

Inject PBS into the filter and pull up the piston of syringe B in order to collect the retentate sample.

Repeat the recovery step by injecting the air in filter, instead of PBS, in order to maximize the recovery of remaining sample.



- Rotate the valve 3 to the position indicated by arrows and disconnect the syringe B.
 - Transfer the retentate sample in a clean tube.

5- Recovery of small EVs.

Collected permeate containing small EVs, soluble proteins and other small molecules can be further concentrated using AMSBIO TFF-Simple(5nm) filter.

Concentrated sample can be further purified with AMSBIO ExoUltra Size Exclusion Chromatography column for downstream application and analyses.

Washing procedure.

Once the filtration process is ended the filter cartridge has to be washed with aboundant MilliQ water. If the cartridge is used for concentrating complex fluids it is recommended to wash as described to point 1, contamination removal.

Procedure for MilliQ water washing

Use 2 clean syringes and load one with MilliQ water (~10ml).

Connect the syringe containing water to the position 5 and the empty one to the position 2.

Close both permeate nozzles (4a and 4b) and rotate the valves 3 and 6 in order to enable the flow through the device (figure 7).

Inject the water in the device and perform washing by pushing alternatively the two syringes upwards and downwards for couple of times.

In final step, push the water in syringe B (position 2) and discard it by detaching the syringe.

Repeat the washing step at least 3 times.

Next, perform washing with MiliQ water while the permeate nozzle 4a is open.

Connect the syringe containing water (10 - 15ml) to the position 5 and the empty one to the position 2.

Flush the filter with water until syringes are completely empty.

Repeat this washing step at least 3 times.

Let the device dry at room temperature.

1- Contamination removal

A- Wash the filter with warm MilliQ water (40-50°C), carefully, applying low pressure with syringes.

B- Wash the filter with a solution of NaOH 0.5 N, then wash with at least 3 volumes of MilliQ water.

2- Optional : washing with Ethanol 96% (5 - 10 ml) can be performed for the removal of potential pyrogenic agents. Wash finally with at lest 3 volumes of MilliQ water.

AMSBIO| www.amsbio.com | info@amsbio.com

UK & Rest of the World 184 Park Drive, Milton Park Abingdon OX 14 4SE. T: +44 (0) 1235 828 200 F: +44 (0) 1235 820 482

North America 1035 Cambridge Street, Cambridge, MA 02141. T: +1 (617) 945-5033 or T: +1 (800) 987-0985 F: +1 (617) 945-8218 Europe Berenkoog 41, 1822 BH Alkmaar, Netherlands T: +31 (0)72 8080244 F: +31 (0)72 8080142 Switzerland Via Lisano 3, (CP.683) CH-6900 T: +41 (0) 91 604 55 22 F: +41 (0) 91 605 17 85 Germany Bockenheimer Landstr. 17/19 60325 Frankfurt/Main T: +49 (0) 69 779099 F: +49 (0) 69 13376880